

ImmunoGuide®

Instructions for Use

Golimumab ELISA

Enzyme immunoassay for the quantitative determination
of free Golimumab in serum and plasma

REF: IG-AA107



12X8



2-8°C



IVD

🏢 TANI Medikal Ltd. Sti.
Turan Gunes Bulvari Korman Sitesi 51/J 06550 Ankara
Tel : +90 312 440 90 20
Fax : +90 312 440 35 15
e-mail: info@immunoguide.com
www.immunoguide.com

<u>Contents</u>	<u>Page</u>
1. Intended Use	2
2. Summary and Explanation	2
3. Principle of the Test	2
4. Warnings and Precautions	3
5. Storage and Stability of the Kit	3
6. Specimen Collection, Handling and Storage	3
7. Contents of the Kit	4
8. Materials Required but not Supplied	4
9. Procedure Notes	5
10. Pre-Test Setup Instructions	5
10.1. Preparation of Components	5
10.2. Dilution of Standards and Samples	6
11. Test Procedure	6
11.1. General Remarks	6
11.2. Assay Procedure	7
11.3. Quality Control	7
11.4. Calculation of Results	8
12. Assay characteristics	9
12.1. Specificity	9
12.2. Sensitivity	9
12.3. Precision	9
12.4. Recovery	9
13. Automation	9
14. References	10

1. INTENDED USE

Enzyme immunoassay for the quantitative determination of free Golimumab in serum and plasma.

2. SUMMARY AND EXPLANATION

The drug Golimumab (trade name Simponi®) is a human monoclonal antibody that binds to both the soluble and transmembrane bioactive forms of human TNF- α . This interaction prevents the binding of TNF- α to its receptors, thereby inhibiting the biological activity of TNF α . Golimumab has been proven effective in the treatment of Rheumatoid Arthritis (RA), Ankylosing Spondylitis (AS), Psoriatic Arthritis (PsA) or Ulcerative Colitis (UC). Antibodies to Golimumab were detected in 57 (4%) of Golimumab-treated patients across the Phase 3 RA, PsA, and AS trials through Week 24. The data from the literature demonstrated that Anti-Drug Antibody positivity was significantly associated with low Golimumab levels and poor therapeutic response. The positive correlation between serum drug trough levels and therapeutic response indicates that drug monitoring could be useful for optimising the dosing of biologics in a personalised therapy strategy. According to the data reported in the prescribing information and the literature, a subcutaneous (SC) injection of 50 mg Golimumab to healthy subjects produced a mean maximum serum concentration (C_{max}) of approximately 2.5 $\mu\text{g/mL}$. When 50 mg Golimumab was administered SC to patients with RA, PsA or AS every 4 weeks, serum concentrations appeared to reach steady state by Week 12. With concomitant use of methotrexate (MTX), treatment with 50 mg Golimumab SC every 4 weeks resulted in a mean steady-state trough serum concentration of approximately 0.4-0.6 $\mu\text{g/mL}$ in patients with active RA, approximately 0.5 $\mu\text{g/mL}$ in patients with active PsA, and approximately 0.8 $\mu\text{g/mL}$ in patients with active AS. In this context, identification of biomarkers for (non-)response and risk factors for adverse drug reactions that might be related to serum concentrations and maintaining the effective concentration of Golimumab in order to potentially avoid some side effects with the reliable method might be beneficial

3. PRINCIPLE OF THE TEST

This ELISA is based on sandwich type ELISA. Diluted standards and samples (serum or plasma) are incubated in the microtiter plate coated with recombinant human TNF- α (rh TNF- α). After incubation, the wells are washed. A horseradish peroxidase (HRP) conjugated anti-human IgG monoclonal antibody is added and binds to the Fc part of Golimumab pre-captured by the rhTNF- α on the surface of the wells. Following incubation, the wells are washed and the bound enzymatic activity is detected by addition of chromogen-substrate. The color developed is proportional to the amount of free Golimumab in the sample or standard. Results of samples can be determined by using the standard curve.

4. WARNINGS AND PRECAUTIONS

1. Before starting the assay, read the instructions completely and carefully. Use the valid version of the package insert provided with the kit. Be sure that everything is understood. For further information (clinical background, test performance, automation protocols, alternative applications, literature, etc.) please refer to the local distributor.
2. In case of severe damage of the kit package, please contact *Tani Medikal* or your supplier in writing, latest one week after receiving the kit. Do not use damaged components in test runs, but keep safe for complaint related issues.
3. Obey lot number and expiry date. Do not mix reagents of different lots. Do not use expired reagents.
4. Follow good laboratory practice and safety guidelines. Wear lab coats, disposable latex gloves and protective glasses where necessary.
5. Reagents of this kit containing hazardous material may cause eye and skin irritations. See MATERIALS SUPPLIED and labels for details.
6. Chemicals and prepared or used reagents have to be treated as hazardous waste according the national biohazard safety guidelines or regulations.
7. Avoid contact with Stop solution. It may cause skin irritations and burns.
8. If any component of this kit contains human serum or plasma it is indicated and if so, it have been tested and were found to be negative for HIV I/II, HBsAg and HCV. However, the presence of these or other infectious agents cannot be excluded absolutely and therefore reagents should be treated as potential biohazards in use and for disposal.
9. Some reagents contain sodium azide (NaN_3) as preservatives. In case of contact with eyes or skin, flush immediately with water. NaN_3 may react with lead and copper plumbing to form explosive metal azides. When disposing reagents, flush with large volume of water to avoid azide build-up

5. STORAGE AND STABILITY OF THE KIT

The kit is shipped at ambient temperature and should be stored at 2-8°C. Keep away from heat or direct sun light. The storage and stability of specimen and prepared reagents is stated in the corresponding chapters. The microtiter strips are stable up to the expiry date of the kit in the broken, but tightly closed bag when stored at 2-8°C.

6. SPECIMEN COLLECTION, HANDLING AND STORAGE

Serum, Plasma (EDTA, Heparin)

The usual precautions for venipuncture should be observed. It is important to preserve the chemical integrity of a blood specimen from the moment it is collected until it is assayed. Do not use grossly hemolytic, icteric or grossly lipemic specimens. Samples appearing turbid should be centrifuged before testing to remove any particulate material.

Storage:	2-8°C	≤-20°C (Aliquots)	Keep away from heat or direct sun light Avoid repeated freeze-thaw cycles
Stability:	3 d	6 mon	

7. CONTENTS OF THE KIT

QUANTITY	COMPONENT
1 x 12 x 8	Microtiter ELISA Plate Break apart strips coated with recombinant human TNF- α (rhTNF- α)
5 x 0.5 mL	Golimumab Standards A-E, Concentrate (10X) 2000; 600; 200; 60; and 0 ng/mL Used for construction of the standard curve. Contains Golimumab, proteins, stabilizer and <15 mM NaN ₃ .
1 x 12 mL	Assay Buffer Blue colored. Ready to use. Contains proteins and <15 mM NaN ₃ .
1 x 60 mL	Dilution Buffer, Concentrate (5X) Contains proteins and <15 mM NaN ₃ .
1 x 12 mL	Enzyme Conjugate Red colored. Ready to use. Contains horseradish peroxidase(HRP)-conjugated anti-human IgG mouse monoclonal antibody, Proclin [®] and stabilizers.
1 x 12 mL	TMB Substrate Solution Ready to use. Contains 3,3',5,5'-Tetramethylbenzidine (TMB).
1 x 12 mL	Stop Solution Ready to use. 1 N Hydrochloric acid (HCl).
1 x 50 mL	Wash Buffer, Concentrate (20x) Contains buffer, Tween [®] 20 and Kathon [™] .
2 x 1	Adhesive Seal For sealing microtiter plate during incubation.

8. MATERIALS REQUIRED BUT NOT SUPPLIED

1. Micropipettes (< 3% CV) and tips to deliver 5-1000 μ L.
2. Bidistilled or deionised water and calibrated glasswares (e.g. flasks or cylinders).
3. Wash bottle, automated or semi-automated microtiter plate washing system.
4. Microtiter plate reader capable of reading absorbance at 450 nm (reference wavelength at 600-650 nm is optional).
5. Absorbent paper towels, standard laboratory glass or plastic vials, and a timer.

9. PROCEDURE NOTES

1. Any improper handling of samples or modification of the test procedure may influence the results. The indicated pipetting volumes, incubation times, temperatures and pre-treatment steps have to be performed strictly according to the instructions. Use calibrated pipettes and devices only.
2. Once the test has been started, all steps should be completed without interruption. Make sure that required reagents, materials and devices are prepared readily at the appropriate time. Allow all reagents and specimens to reach room temperature (20-25 °C) and gently swirl each vial of liquid reagent and sample before use. Mix reagents without foaming.
3. Avoid contamination of reagents, pipettes and wells/tubes. Use new disposable plastic pipette tips for each reagent, standard or specimen. Do not interchange the caps of vials. Always cap not used vials. Do not reuse wells or reagents.
4. Use a pipetting scheme to verify an appropriate plate layout.
5. Incubation time affects results. All wells should be handled in the same order and time sequences. It is recommended to use an 8-channel Micropipettor for pipetting of solutions in all wells.
6. Microplate washing is important. Improperly washed wells will give erroneous results. It is recommended to use a multichannel pipette or an automatic microplate washing system. Do not allow the wells to dry between incubations. Do not scratch coated wells during rinsing and aspiration. Rinse and fill all reagents with care. While rinsing, check that all wells are filled precisely with Wash Buffer, and that there are no residues in the wells.
7. Humidity affects the coated wells. Do not open the pouch until it reaches room temperature. Unused wells should be returned immediately to the resealed pouch including the desiccant.

10. PRE-TEST SETUP INSTRUCTIONS

10.1. Preparation of Components*


Dilute/ dissolve	Component		Diluent	Relation	Remarks	Storage	Stability
10 mL	Wash Buffer	up to 200 mL	Distilled Water	1:20	Warm up at 37°C to dissolve crystals. Mix vigorously.	2-8 °C	4 w
10 mL	Dilution Buffer	up to 50 mL	Distilled Water	1:5		2-8 °C	4 w

* Prepare Wash and Dilution Buffers before starting the assay procedure.

10.2. Dilution of Standards and Samples

The dilutions depicted below are examples of how to obtain final dilutions for standards and samples. Standards and samples should be properly diluted as homogenous mixture before starting the assay procedure. It is recommended mixing the standards and samples well to homogenous solution at each dilution step.

1. 10 μL of standard added to 90 μL of 1X dilution buffer, giving the total volume of 100 μL and a dilution of 1:10.
2. 10 μL of sample added to 490 μL of 1X dilution buffer, giving the total volume of 500 μL and a dilution of 1:50.
3. Samples with a drug concentration above the measuring range should be rated as ">highest standard". The result should not be extrapolated. The sample in question should be further diluted with 1X Dilution Buffer and then retested.



11. TEST PROCEDURE

11.1. GENERAL REMARKS

- 11.1.1. Before performing the assay, samples and assay kit should be brought to room temperature (about 30 minutes beforehand) and ensure the homogeneity of the solution.
- 11.1.2. All Standards should be run with each series of unknown samples.
- 11.1.3. Standards should be subject to the same manipulations and incubation times as the samples being tested.
- 11.1.4. All steps of the test should be completed without interruption.
- 11.1.5. Use new disposable plastic pipette tips for each reagent, standard or specimen in order to avoid cross contamination.

11.2. ASSAY PROCEDURE

1.	Pipette 100μl of Assay Buffer into each of the wells to be used.
2.	Pipette 75 μL of each 1:10 Diluted Standard, and 1:50 Diluted Samples into the respective wells of the microtiter plate. Bubble formation during the pipetting of standards and samples must be avoided. <u>Wells</u> A1: Standard A B1: Standard B C1: Standard C D1: Standard D E1: Standard E F1 and so on: Samples (Serum/Plasma)
3.	Cover the plate with adhesive seal. Shake plate carefully by tapping several times. Incubate the plate on bench top for 60 min at room temperature (RT, 20-25°C).
4.	Remove adhesive seal. Aspirate or decant the incubation solution. Wash the plate 3 X 300 μL of Diluted Wash Buffer per well. Remove excess solution by tapping the inverted plate on a paper towel.
5.	Pipette 100 μL of Enzyme Conjugate (HRP-anti human IgG mAb) into each well.
6.	Cover plate with adhesive seal. Shake plate carefully by tapping several times. Incubate the plate on a bench top for 30 min at RT.
7.	Remove adhesive seal. Aspirate or decant the incubation solution. Wash the plate 3 X 300 μL of Diluted Wash Buffer per well. Remove excess solution by tapping the inverted plate on a paper towel.
8.	Pipette 100 μL of Ready-to-Use TMB Substrate Solution into each well.
9.	Incubate 15 min at RT. Avoid exposure to direct sunlight.
10.	Stop the substrate reaction by adding 100 μL of Stop Solution into each well. Briefly mix contents by gently shaking the plate. Color changes from blue to yellow.
11.	Measure optical density (OD) with a photometer at 450 nm (Reference at OD620 nm is optional) within 15 min after pipetting the Stop Solution.

11.3. QUALITY CONTROL

The test results are only valid if the test has been performed following the instructions. Moreover the user must strictly adhere to the rules of GLP (Good Laboratory Practice) or other applicable standards/laws. All standards/controls must be found within the acceptable ranges. If the criteria are not met, the run is not valid and should be repeated. In case of any deviation, the following technical issues should be reviewed: Expiration dates of (prepared) reagents, storage conditions, pipettes, devices, incubation conditions and washing methods.

11. 4. CALCULATION OF RESULTS


A standard curve should be calculated using the standard concentration (X-axis) versus the OD450 (or OD450/620) values (Y-axis). This can be done manually using graph paper or with a computer program. Concerning the data regression by computer, it is recommended to primarily use the “4 Parameter Logistic (4PL)” or alternatively the “point-to-point calculation”. In case of manual plot there are 2 options: Semilog graph (see Fig. A) or linear graph (see Fig. B). Semilog graph paper is available at <http://www.papersnake.com/logarithmic/semilogarithmic/>.

The concentration of the samples can be read from this standard curve as follows. Using the absorbance value for each sample, determine the corresponding concentration of the drug from the standard curve. This value always has to be multiplied by the individual dilution factor (usually 50). If any diluted sample is reading greater than the highest standard, it should be further diluted appropriately with 1X Dilution Buffer and retested. Also this second dilution has to be used for calculation the final result.

Typical Calibration Curve

(All steps were performed at 23°C. Just an example. Do not use it for calculation!)

1:10 Diluted Standard	A	B	C	D	E
Concentration (ng/mL)	200	60	20	6	0
Mean OD450/620 nm	2.811	1.049	0.396	0.135	0.047



12. ASSAY CHARACTERISTICS

12.1. SPECIFICITY

There is no cross reaction with any other proteins present in naïve serum. Because the solid phase is coated with rhTNF- α , other therapeutic anti-TNF antibodies cause full cross reaction. However, a quantification of other therapeutic antibodies is possible only by using the drug-specific standards, which are available as separate kits from *ImmunoGuide*.

12.2. SENSITIVITY

The lowest detectable level that can be clearly distinguished from the zero standard is 5 ng/mL (zero standard +2SD read from the curve) under the above-described conditions.

12.3. PRECISION

Intra-assay CV: <10%.

Inter-assay CV: <10%.

12.4. RECOVERY

Recovery rate was found to be >95% with native serum and plasma samples when spiked with exogenous Golimumab.

13. AUTOMATION

The *ImmunoGuide* Golimumab ELISA is suitable also for being used by an automated ELISA processor.

14. REFERENCES

1. Gelfer G, Perry L, Deodhar A. Golimumab for the treatment of axial spondyloarthritis. *Expert Rev Clin Immunol*. 2015; 2:1-10.
2. Lichtenstein GR, Hanauer SB, Sandborn WJ. Emerging Treatment Options in Mild to Moderate Ulcerative Colitis. *Gastroenterol Hepatol (N Y)*. 2015;11(3 Suppl 1):1-16.
3. Emery P, Fleischmann RM, Strusberg I, Durez P, Nash P, Amante E, Churchill M, Park W, Pons-Estel B, Han C, Gathany TA, Xu S, Zhou Y, Leu JH, Hsia EC. Efficacy and safety of subcutaneous golimumab in methotrexate-naïve patients with rheumatoid arthritis: 5-year results of the GO-BEFORE trial. *Arthritis Care Res (Hoboken)*. 2015 Oct 16. doi: 10.1002/acr.22759
4. Lee WK, Kim GW, Cho HH, Kim WJ, Mun JH, Song M, Kim HS, Ko HC, Kim MB, Kim BS. Erythrodermic Psoriasis Treated with Golimumab: A Case Report. *Ann Dermatol*. 2015 Aug;27(4):446-9.
5. Brady BL, Tkacz JP, Lofland J, Meyer R, Bolge SC. Prescribing Patterns of Intravenous Golimumab for Rheumatoid Arthritis. *Clin Ther*. 2015;37(9):2028-36.
6. Yonemoto Y, Okamura K, Takeuchi K, Ayabe K, Kaneko T, Matsushita M, Tamura Y, Iso T, Okura C, Otsuka K, Inoue H, Takagishi K. Comparison of golimumab 100-mg monotherapy to golimumab 50 mg plus methotrexate in patients with rheumatoid arthritis: Results from a multicenter, cohort study. *Mod Rheumatol*. 2015; 3:1-5.
7. Abreu MT Commentary: Golimumab in Moderate-to-Severe Ulcerative Colitis. *Gastroenterol Hepatol (N Y)*. 2014;10(7):455-6.
8. Sandborn WJ, Feagan BG, Marano C, Zhang H, Strauss R, Johanns J, Adedokun OJ, Guzzo C, Colombel JF, Reinisch W, Gibson PR, Collins J, Järnerot G, Hibi T, Rutgeerts P; PURSUIT-SC Study Group. Subcutaneous golimumab induces clinical response and remission in patients with moderate-to-severe ulcerative colitis. *Gastroenterology*. 2014;146(1):85-95.
9. Thorlund K, Druyts E, Toor K, Mills EJ. Comparative efficacy of golimumab, infliximab, and adalimumab for moderately to severely active ulcerative colitis: a network meta-analysis accounting for differences in trial designs. *Expert Rev Gastroenterol Hepatol*. 2015;9(5):693-700.
10. Kanbe K, Chiba J, Inoue Y, Taguchi M, Yabuki A. Predictive factors related to the efficacy of golimumab in patients with rheumatoid arthritis. *Clin Med Insights Arthritis Musculoskelet Disord*. 2015;8:25-32.
11. McCarty DJ. Golimumab intravenous infusion for the treatment of rheumatoid arthritis. *Nurse Pract*. 2015;40(3):15-8.
12. Smolen JS, Kay J, Doyle M, Landewé R, Matteson EL, Gaylis N, Wollenhaupt J, Murphy FT, Xu S, Zhou Y, Hsia EC. Golimumab in patients with active rheumatoid arthritis after treatment with tumor necrosis factor α inhibitors: findings with up to five years of treatment in the multicenter, randomized, double-blind, placebo-controlled, phase 3 GO-AFTER study. *Arthritis Res Ther*. 2015 Jan 22;17:14. doi: 10.1186/s13075-015-0516-6.
13. Sato E, Tanaka E, Nakajima A, Inoue E, Shimizu Y, Yamaguchi R, Ochiai M, Shidara K, Hoshi D, Sugimoto N, Seto Y, Taniguchi A, Momohara S, Yamanaka H. Assessment of the effectiveness of golimumab 50-mg and 100-mg regimens in

patients with rheumatoid arthritis in daily practice. *Mod Rheumatol.* 2015;25(4):528-33.

14. Selmi C, Ceribelli A, Naguwa SM, Cantarini L, Shoenfeld Y. Safety issues and concerns of new immunomodulators in rheumatology. *Expert Opin Drug Saf.* 2015;14(3):389-99.

15. Doyle MK, Rahman MU, Frederick B, Birbara CA, de Vries D, Toedter G, Wu X, Chen D, Ranganath VK, Westerman ME, Furst DE. Effects of subcutaneous and intravenous golimumab on inflammatory biomarkers in patients with rheumatoid arthritis: results of a phase 1, randomized, open-label trial. *Rheumatology (Oxford).* 2013;52(7):1214-9.

16. Chen DY, Chen YM, Hung WT, Chen HH, Hsieh CW, Chen YH, Huang WN, Hsieh TY. Immunogenicity, drug trough levels and therapeutic response in patients with rheumatoid arthritis or ankylosing spondylitis after 24-week golimumab treatment. *Ann Rheum Dis.* 2015 Oct 6. pii: annrheumdis-2015-207978. doi: 10.1136/annrheumdis-2015-207978.

17. Huynh C, Reguiai Z, Lambrecht I, Brochot P, Eschard JP, Salmon JH. First two cases of alopecia areata during golimumab therapy. *Joint Bone Spine.* 2015 Oct 7. pii: S1297-319X(15)00178-5. doi: 10.1016/j.jbspin.2015.08.007. [Epub ahead of print]

18. Brunasso AM, Aberer W, Massone C. Subacute lupus erythematosus during treatment with golimumab for seronegative rheumatoid arthritis. *Lupus.* 2014;23(2):201-3.

19. Zhuang Y, Lyn S, Lv Y, Xu Z, Bouman-Thio E, Masterson T, Ford JA, Keen M, Petty KJ, Davis HM, Zhou H. Pharmacokinetics and safety of golimumab in healthy Chinese subjects following a single subcutaneous administration in a randomized phase I trial. *Clin Drug Investig.* 2013;33(11):795-800.

20. Murdaca G, Spanò F, Puppo F. Selective TNF- α inhibitor-induced injection site reactions. *Expert Opin Drug Saf.* 2013;12(2):187-93.